

Ultrasound transducer parameters



Diameter of transducer's measuring surface - 5 mm

Diameter of measuring beam - 2 mm

The angle between measuring surface and transducer holder - 45 degrees

Measurement limits – 0-10 mm, precision 0, 1 mm

There is a measurement fixing button on transducer's handle.